U.S. DEPARTMENT OF AGRICULTURE
FOOD SAFETY AND INSPECTION SERVICE
INTERNATIONAL PROGRAMS

FOREIGN COUNTRY LABORATORY REVIEW

REVIEW DATE	NAME OF FOREIGN LABORATORY
1/21/02	Croatia Veterinary Institute

FOREIGN GOV'T AGENCY	CITY & COUNTRY	ADDRESS OF LABORATORY
	Zagreb, Croatia	Zagreb, Croatia

NAME OF REVIEWER	NAME OF FOREIGN OFFICIAL
Dr. Judd Giezentanner	Dr. Iavorin Bazulic

	Residue Code/Name ▶			Sul	Chl	Clen	Org	Chlo	Diox	Anti	Hor	203			
	REVIEW ITEMS	ITEM #													
SAMPLING PROCEDURES	Sample Handling	01	EVALUATION CODE	A	A	A	A	A	A	A	A	A			
	Sampling Frequency	02		A	A	A	A	A	A	A	A	A			
	Timely Analyses	03		A	A	A	A	A	A	A	A	A			
	Compositing Procedure	04		O	O	O	O	O	O	O	O	O			
	Interpret Comp Data	05		O	O	O	O	O	O	O	O	O			
	Data Reporting	06		A	A	A	A	A	A	A	A	A			
ANALYTICAL PROCEDURES	Acceptable Method	07	EVALUATION CODE	A	A	A	A	A	A	A	A	A			
	Correct Tissue(s)	08		A	A	A	A	A	A	A	A	A			
	Equipment Operation	09		A	A	A	A	A	A	A	A	A			
	Instrument Printouts	10		A	A	A	A	A	A	A	A	A			
QUALITY ASSURANCE PROCEDURES	Minimum Detection Levels	11	EVALUATION CODE	A	A	A	A	A	A	A	A	A			
	Recovery Frequency	12		A	A	A	A	A	A	A	A	A			
	Percent Recovery	13		A	A	A	A	A	A	A	A	A			
	Check Sample Frequency	14		C	C	C	C	C	C	C	C	A			
	All analyst w/Check Samples	15		C	C	C	C	C	C	C	C	A			
	Corrective Actions	16		A	A	A	A	A	A	A	A	A			
	International Check Samples	17		C	C	C	C	C	C	C	C	A			
REVIEW PROCEDURES	Corrected Prior Deficiencies	18	EVAL. CODE	A	A	A	A	A	A	A	A	A			
OTHER REVIEW		19	EVAL. CODE												
		20													

SIGNATURE OF REVIEWER	DATE

FSIS FORM 9500-4 (9/96)

Designed on FormFlow Software

FOREIGN COUNTRY LABORATORY REVIEW *(Comment Sheet)*	REVIEW DATE	NAME OF FOREIGN LABORATORY
	1/21/02	Croatia Veterinary Institute

FOREIGN GOV'T AGENCY	CITY & COUNTRY	ADDRESS OF LABORATORY
	Zagreb, Croatia	Zagreb, Croatia

NAME OF REVIEWER	NAME OF FOREIGN OFFICIAL
Dr. Judd Giezentanner	Dr. Iavorin Bazulic

RESIDUE CODES	ITEM NO.	COMMENTS
All	14/17	Interlaboratory check sample was not performed except for chloramphenicol.
except		
203	15	The interlaboratory check samples were not properly analyzed by supervisor.

U.S. DEPARTMENT OF AGRICULTURE
FOOD SAFETY AND INSPECTION SERVICE
INTERNATIONAL PROGRAMS

FOREIGN COUNTRY LABORATORY REVIEW

REVIEW DATE	NAME OF FOREIGN LABORATORY
1/23/02	Veterinary Institute Rijeka

FOREIGN GOV'T AGENCY	CITY & COUNTRY	ADDRESS OF LABORATORY
	Rijeka, Croatia	Rijeka, Croatia

NAME OF REVIEWER	NAME OF FOREIGN OFFICIAL
Dr. Judd Giezentanner	Dr. S. Perkovic NO U.S. SAMPLES PROCESSED AT THIS LABORATORY

Residue Code/Name ▶

	REVIEW ITEMS	ITEM #	EVALUATION CODE	
SAMPLING PROCEDURES	Sample Handling	01		A
	Sampling Frequency	02		A
	Timely Analyses	03		A
	Compositing Procedure	04		O
	Interpret Comp Data	05		O
	Data Reporting	06		A
ANALYTICAL PROCEDURES	Acceptable Method	07		A
	Correct Tissue(s)	08		A
	Equipment Operation	09		A
	Instrument Printouts	10		A
QUALITY ASSURANCE PROCEDURES	Minimum Detection Levels	11		A
	Recovery Frequency	12		A
	Percent Recovery	13		A
	Check Sample Frequency	14		C
	All analyst w/Check Samples	15		C
	Corrective Actions	16		A
	International Check Samples	17		C
REVIEW PROCEDURES	Corrected Prior Deficiencies	18	EVAL. CODE	A
OTHER REVIEW		19	EVAL. CODE	
		20		

SIGNATURE OF REVIEWER	DATE

FSIS FORM 2520-4 (8/96) Designed on FormFlow Software

FOREIGN COUNTRY LABORATORY REVIEW *(Comment Sheet)*	REVIEW DATE 1/23/02	NAME OF FOREIGN LABORATORY Veterinary Institute Rijeka
FOREIGN GOV'T AGENCY	CITY & COUNTRY Rijeka, Croatia	ADDRESS OF LABORATORY Rijeka, Croatia
NAME OF REVIEWER Dr. Judd Giezentanner	NAME OF FOREIGN OFFICIAL Dr. S. Perkovic	NO U.S. SAMPLES PROCESSED AT THIS LABORATORY

RESIDUE CODES	ITEM NO.	COMMENTS
	14/17	Interlaboratory check sample was not performed except for chloramphenicol.
	15	The interlaboratory check samples were not properly analyzed by supervisor.